

Technical data sheet

miniRITE T

60 85 100 105



	Oticon Opn S 1	Oticon Opn S 2	Oticon Opn S 3	
Speech Understanding	OpenSound Navigator™	Level 1	Level 2	Level 3
	- Balancing power effect	100%	50%	50%
	- Max. noise removal	9 dB	5 dB	3 dB
	OpenSound Optimizer™	•	•	•
	Speech Guard™ LX	Level 1	Level 2	Level 3
	Spatial Sound™ LX	4 estimators	2 estimators	2 estimators
	Soft Speech Booster LX	•	•	•
	Speech Rescue™ LX	•	•	•
Sound Quality	Clear Dynamics	•	•	-
	Spatial Noise Management	•	•	-
	Fitting Bandwidth*	10 KHz	8 KHz	8 KHz
	Processing Channels	64	48	48
	Bass Boost (streaming)	•	•	•
Listening Comfort	Transient Noise Management	4 configurations	On/Off	On/Off
	Feedback shield LX	•	•	•
	Wind Noise Management	•	•	•
Personalization & Optimizing Fitting	YouMatic™ LX	3 configurations	2 configurations	1 configuration
	Fitting Bands	16	14	12
	Multiple Directionality Options	•	•	•
	Adaptation Management	•	•	•
	Oticon Firmware Updater	•	•	•
	Fitting Formulas	VAC+, NAL-NL1 + 2, DSL v5.0	VAC+, NAL-NL1 + 2, DSL v5.0	VAC+, NAL-NL1 + 2, DSL v5.0
Connecting to the World	Stereo streaming (2.4 GHz)	•	•	•
	Oticon ON App	•	•	•
	ConnectClip	•	•	•
	Remote Control 3.0	•	•	•
	TV Adapter 3.0	•	•	•
	Phone Adapter 2.0	•	•	•
	Tinnitus SoundSupport™	•	•	•

* Bandwidth accessible for gain adjustments during fitting

Operating conditions

Temperature: +1°C to +40°C
Relative humidity: 5% to 93%, non-condensing

Storage and transportation conditions

Temperature and humidity should not exceed the following limits for extended periods during transportation and storage.
Temperature: -25°C to +60°C
Relative humidity: 5% to 93%, non-condensing


Oticon Opn S™ miniRITE T is a discreet style, based on the popular miniRITE, and features both telecoil and a convenient double push-button for easy volume and program control.

OpenSound Navigator™ helps users to select and understand speech in all types of environments by balancing the sound sources and attenuating noise.

OpenSound Optimizer™ improves users listening experience and comfort by blocking feedback and securing the targeted amplification of sound sources.

TwinLink™ wireless technology combines binaural communication and 2.4 GHz connectivity with stereo streaming directly from digital devices.

Oticon Opn S is built on the powerful Velox S™ platform which has a programmable firmware architecture, supporting future performance updates.










IP68

oticon
PEOPLE FIRST

For information on compatibility, please visit www.oticon.ca/connectivity

Oticon Opn S 1

miniRITE T 60

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
  Mold, Bass & Power dome Open dome Technical information Omnidirectional mode is used unless otherwise stated.		   Acoustic input: 60 dB SPL Magnetic input: 31.6 mA/m	   Acoustic input: 60 dB SPL Magnetic input: 31.6 mA/m
OSPL90	Peak	116 dB SPL	105 dB SPL
	1600 Hz	109 dB SPL	100 dB SPL
	HFA-OSPL90	110 dB SPL	102 dB SPL
Full-on gain ¹	Peak	46 dB	35 dB
	1600 Hz	37 dB	29 dB
	HFA-FOG	38 dB	30 dB
Reference test gain		30 dB	26 dB
Frequency range		110-9700 Hz	100-9200 Hz
Telecoil output (1600 Hz)	1 mA/m field	67 dB SPL	-
	10 mA/m field	87 dB SPL	-
	SPLITS L/R	-	85/85 dB SPL
Total harmonic distortion (Input 70 dB SPL)	500 Hz	<2 %	<2 %
	800 Hz	<3 %	<2 %
	1600 Hz	<2 %	<2 %
Equivalent input noise level	Omni	21 dB SPL	18 dB SPL
	Dir	28 dB SPL	27 dB SPL
Battery consumption ²	Typical	1.5 mA	1.6 mA
	Quiescent	1.5 mA	1.5 mA
Battery life, artificial measurement, hours ³		120	115
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		60-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 16/21/26 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.








2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.

4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 2 & 3

miniRITE T 60

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
  Mold, Bass & Power dome Open dome Technical information Omnidirectional mode is used unless otherwise stated.		   Acoustic input: 60 dB SPL Magnetic input: 31.6 mA/m	   Acoustic input: 60 dB SPL Magnetic input: 31.6 mA/m
OSPL90	Peak	116 dB SPL	105 dB SPL
	1600 Hz	109 dB SPL	100 dB SPL
	HFA-OSPL90	110 dB SPL	102 dB SPL
Full-on gain ¹	Peak	46 dB	35 dB
	1600 Hz	37 dB	29 dB
	HFA-FOG	38 dB	30 dB
Reference test gain		30 dB	26 dB
Frequency range		110-7500 Hz	100-7500 Hz
Telecoil output (1600 Hz)	1 mA/m field	67 dB SPL	-
	10 mA/m field	87 dB SPL	-
	SPLITS L/R	-	85/85 dB SPL
Total harmonic distortion (Input 70 dB SPL)	500 Hz	<2 %	<2 %
	800 Hz	<3 %	<2 %
	1600 Hz	<2 %	<2 %
Equivalent input noise level	Omni	22 dB SPL	19 dB SPL
	Dir	30 dB SPL	28 dB SPL
Battery consumption ²	Typical	1.5 mA	1.6 mA
	Quiescent	1.5 mA	1.5 mA
Battery life, artificial measurement, hours ³		120	115
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		60-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 16/21/26 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.








2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.

4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 1

miniRITE T 85

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
 85 Technical information Omnidirectional mode is used unless otherwise stated.		 — Acoustic input: 60 dB SPL - - - Magnetic input: 31.6 mA/m	 — Acoustic input: 60 dB SPL - - - Magnetic input: 31.6 mA/m
OSPL90	Peak 1600 Hz HFA-OSPL90	127 dB SPL 120 dB SPL 121 dB SPL	116 dB SPL 111 dB SPL 112 dB SPL
Full-on gain ¹	Peak 1600 Hz HFA-FOG	66 dB 52 dB 55 dB	54 dB 43 dB 47 dB
Reference test gain		45 dB	34 dB
Frequency range		120-9500 Hz	100-8500 Hz
Telecoil output (1600 Hz)	1 mA/m field 10 mA/m field SPLITS L/R	82 dB SPL 102 dB SPL -	- - 94/94 dB SPL
Total harmonic distortion (Input 70 dB SPL)	500 Hz 800 Hz 1600 Hz	<2 % <3 % <2 %	<2 % <2 % <2 %
Equivalent input noise level	Omni Dir	25 dB SPL 32 dB SPL	20 dB SPL 29 dB SPL
Battery consumption ²	Typical Quiescent	1.6 mA 1.5 mA	1.7 mA 1.5 mA
Battery life, artificial measurement, hours ³		110	105
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		55-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 20/20/24 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.








2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.

4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 2 & 3

miniRITE T 85

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006	
 85 Technical information Omnidirectional mode is used unless otherwise stated.		 Acoustic input: 60 dB SPL Magnetic input: 31.6 mA/m	 Acoustic input: 60 dB SPL Magnetic input: 31.6 mA/m	
OSPL90		Peak 1600 Hz HFA-OSPL90	127 dB SPL 120 dB SPL 121 dB SPL	116 dB SPL 111 dB SPL 112 dB SPL
Full-on gain ¹		Peak 1600 Hz HFA-FOG	66 dB 52 dB 55 dB	54 dB 43 dB 47 dB
Reference test gain			45 dB	34 dB
Frequency range			120-7500 Hz	100-7500 Hz
Telecoil output (1600 Hz)		1 mA/m field 10 mA/m field SPLITS L/R	82 dB SPL 102 dB SPL -	- - 94/94 dB SPL
Total harmonic distortion (Input 70 dB SPL)		500 Hz 800 Hz 1600 Hz	<2 % <3 % <2 %	<2 % <2 % <2 %
Equivalent input noise level		Omni Dir	26 dB SPL 33 dB SPL	21 dB SPL 30 dB SPL
Battery consumption ²		Typical Quiescent	1.6 mA 1.5 mA	1.7 mA 1.5 mA
Battery life, artificial measurement, hours ³			110	105
Expected battery life, hours (battery size 312 - IEC PR41) ⁴			55-65	
IRIL (IEC 60118-13:2011)			700/1400/2000 MHz: 20/20/24 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.








2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.

4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 1








miniRITE T 100

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
  Power flex mold, Bass & Power dome		  	  
Technical information Omnidirectional mode is used unless otherwise stated.			
Instrument warning The maximum output capability of the hearing instrument may exceed 132 dB SPL (IEC 711). Special care should be exercised in selecting and fitting the instrument as there may be risk of impairing the remaining hearing of the hearing aid user.			
OSPL90		Peak 132 dB SPL 1600 Hz 130 dB SPL HFA-OSPL90 127 dB SPL	Peak 122 dB SPL 1600 Hz 121 dB SPL HFA-OSPL90 118 dB SPL
Full-on gain ¹		Peak 66 dB 1600 Hz 56 dB HFA-FOG 59 dB	Peak 57 dB 1600 Hz 48 dB HFA-FOG 51 dB
Reference test gain		49 dB	42 dB
Frequency range		100-8500 Hz	100-8000 Hz
Telecoil output (1600 Hz)		1 mA/m field 86 dB SPL 10 mA/m field 106 dB SPL SPLITS L/R -	- - 103/103 dB SPL
Total harmonic distortion (Input 70 dB SPL)		500 Hz <7 % 800 Hz <4 % 1600 Hz <2 %	<2 % <2 % <2 %
Equivalent input noise level		Omni 23 dB SPL Dir 32 dB SPL	19 dB SPL 30 dB SPL
Battery consumption ²		Typical 1.5 mA Quiescent 1.5 mA	1.7 mA 1.5 mA
Battery life, artificial measurement, hours ³		115	105
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		50-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 18/21/28 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.
2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.
3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.
4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 2 & 3






miniRITE T 100

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
  Power flex mold, Bass & Power dome		  	  
Technical information Omnidirectional mode is used unless otherwise stated.			
Instrument warning The maximum output capability of the hearing instrument may exceed 132 dB SPL (IEC 711). Special care should be exercised in selecting and fitting the instrument as there may be risk of impairing the remaining hearing of the hearing aid user.			
OSPL90		Peak 132 dB SPL 1600 Hz 130 dB SPL HFA-OSPL90 127 dB SPL	Peak 122 dB SPL 1600 Hz 121 dB SPL HFA-OSPL90 118 dB SPL
Full-on gain ¹		Peak 66 dB 1600 Hz 56 dB HFA-FOG 59 dB	Peak 57 dB 1600 Hz 48 dB HFA-FOG 51 dB
Reference test gain		49 dB	42 dB
Frequency range		100-7500 Hz	100-7500 Hz
Telecoil output (1600 Hz)		1 mA/m field 86 dB SPL 10 mA/m field 106 dB SPL SPLITS L/R -	- - 103/103 dB SPL
Total harmonic distortion (Input 70 dB SPL)		500 Hz <7 % 800 Hz <4 % 1600 Hz <2 %	<2 % <2 % <2 %
Equivalent input noise level		Omni 23 dB SPL Dir 32 dB SPL	19 dB SPL 30 dB SPL
Battery consumption ²		Typical 1.5 mA Quiescent 1.5 mA	1.7 mA 1.5 mA
Battery life, artificial measurement, hours ³		115	105
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		50-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 18/21/28 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.
2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.
3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.
4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 1

miniRITE T 105

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
		  	  
Technical information Omnidirectional mode is used unless otherwise stated.			
Instrument warning The maximum output capability of the hearing instrument may exceed 132 dB SPL (IEC 711). Special care should be exercised in selecting and fitting the instrument as there may be risk of impairing the remaining hearing of the hearing aid user.			
— Acoustic input: 60 dB SPL - - - Magnetic input: 31.6 mA/m			
OSPL90	Peak	135 dB SPL	127 dB SPL
	1600 Hz	132 dB SPL	125 dB SPL
	HFA-OSPL90	130 dB SPL	122 dB SPL
Full-on gain ¹	Peak	72 dB	64 dB
	1600 Hz	65 dB	57 dB
	HFA-FOG	65 dB	57 dB
Reference test gain		58 dB	46 dB
Frequency range		100-8200 Hz	100-7800 Hz
Telecoil output (1600 Hz)	1 mA/m field	96 dB SPL	-
	10 mA/m field	116 dB SPL	-
	SPLITS L/R	-	105/105 dB SPL
Total harmonic distortion (Input 70 dB SPL)	500 Hz	<2 %	<2 %
	800 Hz	<2 %	<2 %
	1600 Hz	<3 %	<2 %
Equivalent input noise level	Omni	18 dB SPL	18 dB SPL
	Dir	28 dB SPL	29 dB SPL
Battery consumption ²	Typical	1.6 mA	1.7 mA
	Quiescent	1.5 mA	1.5 mA
Battery life, artificial measurement, hours ³		110	105
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		45-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 38/18/39 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.








2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.

4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Oticon Opn S 2 & 3

miniRITE T 105

Technical data		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010	ZCC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
		  	  
Technical information Omnidirectional mode is used unless otherwise stated.			
Instrument warning The maximum output capability of the hearing instrument may exceed 132 dB SPL (IEC 711). Special care should be exercised in selecting and fitting the instrument as there may be risk of impairing the remaining hearing of the hearing aid user.			
— Acoustic input: 60 dB SPL - - - Magnetic input: 31.6 mA/m			
OSPL90	Peak	135 dB SPL	127 dB SPL
	1600 Hz	132 dB SPL	125 dB SPL
	HFA-OSPL90	130 dB SPL	122 dB SPL
Full-on gain ¹	Peak	72 dB	64 dB
	1600 Hz	65 dB	57 dB
	HFA-FOG	65 dB	57 dB
Reference test gain		58 dB	46 dB
Frequency range		100-7500 Hz	100-6500 Hz
Telecoil output (1600 Hz)	1 mA/m field	96 dB SPL	-
	10 mA/m field	116 dB SPL	-
	SPLITS L/R	-	105/105 dB SPL
Total harmonic distortion (Input 70 dB SPL)	500 Hz	<2 %	<2 %
	800 Hz	<2 %	<2 %
	1600 Hz	<3 %	<2 %
Equivalent input noise level	Omni	18 dB SPL	18 dB SPL
	Dir	28 dB SPL	29 dB SPL
Battery consumption ²	Typical	1.6 mA	1.7 mA
	Quiescent	1.5 mA	1.5 mA
Battery life, artificial measurement, hours ³		110	105
Expected battery life, hours (battery size 312 - IEC PR41) ⁴		45-65	
IRIL (IEC 60118-13:2011)		700/1400/2000 MHz: 38/18/39 dB SPL	

1) Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.

2) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

3) Based on the standardized battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.

4) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

This image shows a full page of white paper with horizontal grey ruling lines. The word "Notes" is printed at the top left. There are 20 horizontal lines in total, creating 19 equal-sized rows for writing.

[illegible]



Manufactured by:
Oticon A/S
Kongebakken 9
DK-2765 Smørum
Denmark
www.oticon.global

Imported and
Distributed by:
Audmet Canada Ltd
1600-4950 Yonge St
Toronto, ON M2N 6K1
www.oticon.ca

www.oticon.ca

oticon
PEOPLE FIRST